

CONFIDENTIAL

903304

NPIC/TDS/D/6-1729  
20 December 1966

MEMORANDUM FOR THE RECORD

SUBJECT: Contract [REDACTED]

25X1A 1. Upon further inspection of the light tables developed under  
25X1A the subject contract, certain deficiencies were discovered with  
25X1A regard to the microscope adapters.

25X1A 2. Specifically, the adapter for the [REDACTED] Zoom 70 for light  
25X1A table #2 could not be collimated. [REDACTED] had not collimated it  
25X1A to the degree NPIC requires (10 minutes of arc) nor had they allowed  
25X1A for the adjustment. Since it was undesirable to send the light  
25X1A tables back to the contractor [REDACTED] and the undersigned  
25X1A modified the adapter by disassembling it and enlarging the mounting  
25X1A holes between the mounting block and the microscope ring. This  
25X1A effort required about five man hours.

25X1A 3. The [REDACTED] adapter for light table #3 had been assembled  
25X1A incorrectly by [REDACTED] in that the focusing mechanism gear had  
25X1A been assembled backwards and serious damage to both the gear and its  
25X1A rack had occurred. This item is being returned to [REDACTED] for re-work.

[REDACTED]  
Development Staff, TDS

Distribution:

Orig - Project File  
2 - Chonos

pnf

DECLASS REVIEW BY NIMA / DoD

CONFIDENTIAL